

Certificate



Certificate no.

TA 50430169 01

License Holder:

MEAN WELL Enterprises Co., Ltd.
No.28, Wuquan 3rd Rd.,
Wugu District, New Taipei City
24891
Taiwan

Manufacturing Plant:

Refer to latest revision
of the annex list of factories

Test report no.: ZTW1- 50045257 002

Client Reference: ES-23102018-TUV

Tested to: IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12

Certified Product: (Switching Mode Power Supply)

License Fee - Units

Type Designation	: RPS-400-xy (MW)	10
x stands for	: 12, 15, 18, 24, 27, 36 or 48	1
y stands for	: blank, -C, -TF or -SF	1
Rated Voltage	: AC 100-240V, 50/60Hz	
Rated Current	: 4.8-2.1A	
Outputs	: see constructional dataform	
max. Ambient Temperature	: see constructional dataform	
max. Operational Altitude	: 4000m	
Protection Class	: I	

Remarks: 1) The risk management requirements of the standard were not addressed;
2) This certificate is valid for maximum 5 years.

Appendix: 1

12

Licensed Test mark:



TÜV Rheinland Taiwan Ltd

Signature

Date of Issue

(day/mo/yr)

27/02/2019



Dipl.-Ing. (FH) A. Klinker

**Anlage Fertigungsstättenliste
/Attachment List of Factories**



TA 50430169 0001

1 MEAN WELL Enterprises Co., Ltd.
No.28, Wuquan 3rd Rd.,
Wugu District, New Taipei City
24891
Taiwan

2 SuZhou MEAN WELL Technology
Co., Ltd.
No. 77, Jian-min Road, Dong-qiao, Pan-
yang Ind. Park, Huang-dai Town, Xiang-
cheng District Suzhou, Jiangsu 215152
P. R. China

3 MEAN WELL (GUANGZHOU) ELECTRONICS
CO., LTD HUADU BRANCH
No. 11 Jingu South Road,
Huadong Town, Huadu District,
Guangzhou 510890
P. R. China



Dieser Anhang ersetzt den Vorgänger vom/
This annex replaces the previous annex dated

Datum / Date 27.02.2019



Zertifizierungsstelle
/ Certification Body

A handwritten signature in blue ink, appearing to be "A. Klinker".

Dipl.-Ing. (FH) A. Klinker

Certificate



Certificate no.

TA 50430169 02

License Holder:

MEAN WELL Enterprises Co., Ltd.
No.28, Wuquan 3rd Rd.,
Wugu District, New Taipei City
24891
Taiwan

Manufacturing Plant:

MEAN WELL Enterprises Co., Ltd.
No.28, Wuquan 3rd Rd.,
Wugu District, New Taipei City
24891
Taiwan

Test report no.: ZTW1- 50045257 002

Client Reference: ES-23102018-TUV

Tested to: IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12

Certified Product: (Switching Mode Power Supply)

License Fee - Units

as page 01

Addition

Type Designation	: RPS-500-xy (MW)	1
x stands for	: 12, 15, 18, 24, 27, 36 or 48	1
y stands for	: blank, -C, -TF or -SF	1
Rated Voltage	: AC 100-240V, 50/60Hz	
Rated Current	: 5.9-2.7A	

This certificate is valid for maximum 5 years.

Appendix: 1

3

Licensed Test mark:



TÜV Rheinland Taiwan Ltd

Signature

Date of Issue

(day/mo/yr)

27/02/2019



Dipl.-Ing. (FH) A. Klinker

Certificate



Certificate no.

TA 50430169 03

License Holder:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu District,
New Taipei City 24891
Taiwan

Manufacturing Plant:

Refer to latest revision
of the annex list of factories

Test report no.: ZTW1-LinPr 50045257 003 **Client Reference:** P00818639

Tested to: IEC 60601-1:2005+A1
EN 60601-1:2006+A1+A12

Certified Product: (Switching Mode Power Supply)

License Fee - Units

as page 01

Change

Test Requirement : see above

This certificate is valid for maximum 5 years.

Appendix: 1

Licensed Test mark:



TÜV Rheinland Taiwan Ltd

Signature

Dipl.-Ing. (FH) A. Klinker



**Date of Issue
(day/mo/yr)**

27/03/2023

Anlage Fertigungsstättenliste
/Attachment List of Factories



TA 50430169 0001

Anlagenrevision / Annex revision

1

1 MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu District,
New Taipei City 24891
Taiwan

2 MEAN WELL (GUANGZHOU) ELECTRONICS
CO., LTD.
No. 11,
Jingu South Road, Huadu District,
510890 Guangzhou,
P.R. China

3 Suzhou MEAN WELL Technology Co., Ltd.
No.269, Changping Road, Huangdai Town,
Xiangcheng District, Suzhou
Jiangsu
P.R. China



Dieser Anhang ersetzt den Vorgänger vom/
This annex replaces the previous annex dated

Datum / Date 27.02.2019



Zertifizierungsstelle
/ Certification Body

Dipl.-Ing. (FH) A. Klinker

Certificate



Certificate no.

TA 50430169 04

License Holder:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu District,
New Taipei City 24891
Taiwan

Manufacturing Plant:

Refer to latest revision
of the annex list of factories

Test report no.: ZTW1-RC 50045257 004

Client Reference: P00984755

Tested to: IEC 60601-1:2005+A1+A2
EN 60601-1:2006+A1+A12+A2

Certified Product: (Switching Mode Power Supply)

License Fee - Units

as page 01

Change

Test Requirement : see above

Remark: 1. The risk management requirements of the standard were not addressed.
2. This certificate is valid for maximum 5 years from last page.

Appendix: 3

Licensed Test mark:




IEC 60601-1
EN 60601-1

www.tuv.com
ID 1419066516

TÜV Rheinland Taiwan Ltd

Signature


Dipl.-Ing. (FH) A. Klinker



Date of Issue
(day/mo/yr)

08/12/2023